

STATEMENT OF CONFORMITY

Project No. 10703

We hereby confirm that the applicant has successfully demonstrated that his product is in conformity with the EMC standards referred below. The tests have been conducted in the Nemko AS EMC facilities in Norway.

Applicant Norsk Elektro Optikk AS
Solheimsveien 62A
Postboks 384
1471 Lørenskog, NORWAY

Manufacturer Norsk Elektro Optikk AS
Solheimsveien 62A
Postboks 384
1471 Lørenskog, NORWAY

Product Gas Monitor

Model/type Lasergas II

Data Max 20W 115-230V AC 50/60Hz or 18-36V DC

Other specification

Brand Name

Standards applied EN 61000-6-2:2001
EN 61000-6-3:2001
EN 61326-1:1997 + A1:1998

Revision Rev. A. Revision due to missing voltage rating and standard.

Remarks

This statement is a proof of compliance given by Nemko after the product has undergone a fully valid EMC tests according to and compliant to the specified standards.

This statement is given instead of an EMC test report. The test documentation necessary to produce a test report is held by Nemko for 10 years. The applicant may at any time order a test report in addition to this statement.

This statement is based on a single evaluation of one sample of the above mentioned product. It does not imply an assessment of the whole production. It is the manufacturer's responsibility to assure the additional production units of this product are manufactured with identical electrical and mechanical components. The manufacturer is responsible to the Competent Authorities in Europe for any modifications made to the product which result in non-compliance to the relevant regulations.

Date of issue 20 June 2003


Tested by



signature

Jon Fredrik Mo

Authorized by



signature

Roger Berget [Head of section]